

Aerobic and Anaerobic Bacteria Test Kit



Brand:

Product Code: 180-50

Availability: Call for availability

Lead Time: 35

Description

The procedure utilizes an extinction dilution procedure to identify and count the number of organisms present. A minimum of five vials and syringes are required for each test with a twenty-eight day incubation period for anaerobic and a minimum five day period for aerobic vials.

Size: 6.5" × 6.5" × 6.5" (17 × 17 × 17 cm)
Weight: 13 oz (368 g)